

## CERTIFICATE OF ANALYSIS

### Bicinchoninic Acid Kit for Protein Determination - Micro

Catalog No.

PRTD2

Lot No.

CM21A-RP

Lot No.

Reagent A

CM21B-RP

Reagent B

CM21C-RP

Reagent C


CM21D-RP

Storage: Room Temperature

Expiry Date: 11/2013

TEST	REQUIREMENTS	RESULTS
Absorbance at 562 nm of freshly prepared Standard Working Reagent	$\leq 0.07$	Pass

Approved:


Quality control  
Technician.

Reviewed:


Dr. P. Jansen  
Production Manager

Date:

21/11/2011