

CERTIFICATE OF ANALYSIS

Product No.

PRTD2

Lot No.

EB22A-GR

QPRO-BCA kit Micro

Comments: Bicinchonic acid kit for protein determination
CAS No: /
Components: Solution A: Tartrate in alkaline buffer (240mL), Lot EB22B-GR
Solution B: 4% BCA in water (240mL), Lot EB22C-GR
Solution C: 4% CuSO4 5H2O in water (12mL), Lot EB22D-GR
Storage: Room Temperature
Expiry Date: 02/2015

TEST	SPECIFICATIONS	LOT RESULT
Absorbance at 562 nm of freshly prepared standard working reagent	Abs \leq 0.07	Pass

Quality Release Date: 26/02/2013

Quality Retest Date: --/--/----

Approved:


Quality Control Technician

Reviewed:


Dr. T. P. Jansen
Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.