

## CERTIFICATE OF ANALYSIS

### Bicinchoninic Acid Kit for Protein Determination - Standard

Catalog No.

**PRTD1**

Lot No.

**CN20A-RP**

Lot No.

**CN20B-RP**

**CN20C-RP**

**Reagent A**

**Reagent B**

**Storage:** Room Temperature

**Expiry Date:** 12/2013

TEST	REQUIREMENTS	RESULTS
Absorbance at 562 nm of freshly prepared Standard Working Reagent	$\leq 0.07$	Pass

Approved:



Quality control  
Technician.

Reviewed:



Dr. R. Perciaccante  
Production Manager

Date:

**20/12/2011**