

CERTIFICATE OF ANALYSIS

Bicinchoninic Acid Kit for Protein Determination - Standard

Catalog No.

PRTD1

Lot No.

DC06A-RP

Lot No.

Reagent A**DC06B-RP****Reagent B****DC06C-RP**

Storage: Room Temperature


Expiry Date: 03/2014

TEST	REQUIREMENTS	RESULTS
Absorbance at 562 nm of freshly prepared Standard Working Reagent	≤ 0.07	Pass

Approved:


Quality control
Technician.

Reviewed:


Dr. R. Perciaccante
Production Manager

Date:

07/03/2012