

CERTIFICATE OF ANALYSIS

Bicinchoninic Acid Kit for Protein Determination - Standard

Catalog No.

PRTD1

Lot No.

DE17A-GR

Lot No.

Reagent A

Reagent B

DE17B-GR

DE17B-GR

Storage: Room Temperature

Expiry Date: 05/2014


TEST	REQUIREMENTS	RESULTS
Absorbance at 562 nm of freshly prepared Standard Working Reagent	≤ 0.07	Pass

Approved:



Quality control
Technician.

Reviewed:



Dr. P. Jansen
Production Manager

Date:

17/05/2012