CYANAGEN
Reagents for Molecular Biology

CYANAGEN SRL
Sede Legale/Registered Site:
Via Andrea Costa 4/2 - 40134 Bologna (BO) - Italy
Sede Operativa/Headquarter:
Via degli Stradelli Guelfi 40/c – 40138 Bologna (BO) - Italy
Tel./fax: +39 051 534063 Web: www.cyanagen.com
e-mail: info@cyanagen.com - PEC: mail@pec.cyanagen.it

CERTIFICATE OF ANALYSIS

STAR BEADS DNA Extraction Kit, Prefilled Plates

Product Information

Product Code:	SBK307,2X32
Lot Number:	TF22A-XP
Description:	Kit ready-to-use designed for the rapid and efficient isolation of high-quality DNA from a wide variety of specimens.
Components:	4x **STAR BEADS DNA Extraction Plate** Cod. SBK262,1X16PFI, Lot TF22B-XP 1x **STAR BEADS Binding Buffer 1** Cod. SBBB283ML, Lot TF22C-XP 2x **STAR BEADS Proteinase k**, Cod. SBK263,0750, Lot 181125
Storage:	15-25°C
Expiry date:	2027/06

TEST	SPECIFICATIONS	LOT RESULT
Positive Control Test:		
dsDNA extraction from 200 µl whole blood and OD/fluorimetric analysis	Conc (ng/µl) ≥ 20 ng/µl A260/A280 ≥ 1,7.	27,51 1,91
Negative Control Test:		
dsDNA extraction from no sample and OD/fluorimetric analysis	Conc (ng/µl) < 1 ng/µl	0,25

Document Release Date: 24/06/2026
Document Revision: 00

Approved: Marina Blumetti
QC Manager

Reviewed: Federica Radeghieri
Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.
MO 7.5.1.1 Rev04