

CERTIFICATE OF ANALYSIS

STAR BEADS Genomic DNA Extraction Kit, Prefilled Plates

Product Information

Product Code:	SBK287,1X96PFI
Lot Number:	QG20A-XP
Description:	Kit ready-to-use designed for the rapid and efficient isolation of Genomic DNA from blood and saliva. CE-IVD
Components:	<p>1x STAR BEADS Genomic Sample Plate, Cod. SBSP288,1X96PFI, Lot QG20B-XP</p> <p>1x STAR BEADS Magnetic Beads Plate, Cod. SBBP293,1X96PFI, Lot QG20C-XP</p> <p>1x STAR BEADS Genomic Binding Plate, Cod. SBBP293,1X96PFI, Lot QG20D-XP</p> <p>2x STAR BEADS Genomic Washing 1 Plate, Cod. SBWP290,1X96PFI, Lot QG20E-XP</p> <p>1x STAR BEADS Genomic Washing 2 Plate, Cod. SBWP291,1X96PFI, Lot QG20F-XP</p> <p>1x STAR BEADS Genomic Elution Plate, Cod. SBEP292,1X96PFI, Lot QG20G-XP</p> <p>1x STAR BEADS Genomic Tip Comb Plate, Cod. SBTP296,1X96PFI, Lot QG20H-XP</p> <p>3x STAR BEADS Proteinase k, Cod. SBK263,0750, Lot 020223</p>
Storage:	15-25°C
Expiry date:	07/2024

TEST	SPECIFICATIONS	LOT RESULT
Positive Control Test: dsDNA extraction from 200 µl whole blood and OD/fluorimetric analysis	Conc (ng/µl) ≥ 20 ng/µl A260/A280 ≥ 1,7.	47.15 ng/µl 1.86
Negative Control Test: dsDNA extraction from no sample and OD/fluorimetric analysis	Conc (ng/µl) < 1 ng/µl	0.53

Document Release Date: 05/10/2023

Document Revision: 00

Approved:

Maria Elena Lattuada

QC Manager

Reviewed:

Federica Rodighiero

Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.

MO 7.5.1.1 Rev04