

## CERTIFICATE OF ANALYSIS

### STAR BEADS Genomic DNA Extraction Kit, Prefilled Plates


#### Product Information

Product Code:	SBK261,2X32PFI
Lot Number:	RC20A-XP
Description:	Kit ready-to-use designed for the rapid and efficient isolation of Genomic DNA from blood and saliva. <b>CE-IVD</b>
Components:	4x <b>STAR BEADS Genomic Extraction Plate</b> , Cod. SBK262,1X16PFI, Lot RC18A-XP 1x <b>STAR BEADS Genomic Binding Buffer</b> , Cod. SBBB283,030, Lot RC20B-XP 2x <b>STAR BEADS Proteinase k</b> , Cod. SBK263,0750, Lot 020223
Storage:	15-25°C
Expiry date:	2025-02


TEST	SPECIFICATIONS	LOT RESULT
<b>Positive Control Test:</b> dsDNA extraction from 200 µl whole blood and OD/fluorimetric analysis	Conc (ng/µl) ≥ 20 ng/µl A260/A280 ≥ 1,7.	32.14 1.79
<b>Negative Control Test:</b> dsDNA extraction from no sample and OD/fluorimetric analysis	Conc (ng/µl) < 1 ng/µl	Conc < 1 ng/µl

Document Release Date: 22/03/2024

Document Revision: 00

Approved: 

QC Manager

Reviewed: 

Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.

MO 7.5.1.1 Rev04