

## CERTIFICATE OF ANALYSIS

### STAR BEADS Universal DNA/RNA Extraction Kit

#### Product Information

<b>Product Code:</b>	SBK273,2x32PFI
<b>Lot Number:</b>	RI09A-XP
<b>Description:</b>	Kit ready-to-use designed for the rapid and efficient isolation of nucleic acids from a wide variety of specimens such as biological fluids, cultured cells, formalin-fixed-paraffin-embedded (FFPE) samples, frozen and fresh tissues, small-size organisms i.e. insects and food.
<b>Components:</b>	4 x STAR BEADS Extraction Plate SBK251,1X16PFI, Lot RF25C-XP 2 x STAR BEADS Proteinase K SBK263,0750, Lot 020223
<b>Storage:</b>	15-25°C
<b>Stability:</b>	06/2025

TEST	SPECIFICATIONS	LOT RESULT
Positive Control Test:	32,34 < Ct (CMV 6 cp) < 37,76	Ct (CMV 6 cp) = 35,46
Negative Control Test:	No amplification	No amplification

Document Release Date: 09/09/2024

Document Revision: 00

Approved:

  
QC Manager

Reviewed:

  
Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.  
MO 7.5.1.1 Rev04