

CERTIFICATE OF ANALYSIS

Product No.

SBK186,1x96PF

Lot No.

OF29F-XP

STAR BEADS Viral DNA/RNA Extraction Kit, prefilled plates

Comments: Kit designed for the rapid and efficient isolation of viral DNA and RNA from serum, swabs, plasma, saliva, and other body fluids

- **STAR BEADS Viral DNA/RNA Sample Plate**

prefilled with STAR BEADS Lysis buffer and STAR BEADS Magnetic Beads, SBSP194,1x96PF Lot OF29G-XP

- **STAR BEADS Viral DNA/RNA Washing 1 Plate**

prefilled with ready to use STAR BEADS Washing Buffer 1 SBWP192,1X96PF, Lot OF29H-XP

Components:

- **STAR BEADS Viral DNA/RNA Washing 2 Plate**

prefilled with ready to use STAR BEADS Washing Buffer 2 SBWP193,1X96PF, Lot OF29I-XP

- **STAR BEADS Viral DNA/RNA Elution Plate**

SBEP195,1X96PF, Lot OF29L-XP

- **STAR BEADS Viral DNA/RNA Tip Comb Plate**

SBTP196,1X96PF, Lot OF29M-XP

Storage: 15-30°C

Expiry Date: 06/2022

| TEST | SPECIFICATIONS | LOT RESULT |
|--------------------------------------|---|------------|
| Nucleic acid extraction and qRT-PCR: | Average CV ≤10.00% (sample Ct vs reference Ct) | 0.93% |

Document Release Date: 05/07/2021

Document Revision: 00

Approved:

Quality Control Technician

Reviewed:

Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.

CYANAGEN SRL -

C.F./P.IVA: 02360111203 iscritta nel Registro Imprese di Bologna n. REA: 433637

Capitale sociale sottoscritto e versato € 10.000,00.

"Società soggetta a direzione e coordinamento da parte di Finance & Technology Srl - art.2497 bis cc".

