

## CERTIFICATE OF ANALYSIS

**Product No.**

**SBK233,1x96PFI**

**Lot No.**

**PA27A-XP**

### STAR BEADS Viral DNA/RNA Extraction kit Prefilled Plates

**Comments:**

Kit designed for the rapid and efficient isolation of viral DNA and RNA from serum, swabs, plasma, saliva, and other body fluids

- **STAR BEADS Viral Sample Plate**, SBSP232,1x96PFI, Lot PA27B-XP  
Prefilled with STAR BEADS Viral Lysis – Binding Buffer

- **STAR BEADS Washing 1 Plate**, SBWP192,1X96PF, Lot PA27C-XP  
Prefilled with STAR BEADS Washing Buffer 1

**Components:**

- **STAR BEADS Washing 2 Plate**, SBWP193,1X96PF, Lot PA27D-XP  
Prefilled with STAR BEADS Washing Buffer 2

- **STAR BEADS Elution Plate**, SBEP230,1X96PFI, Lot PA27E-XP  
Prefilled with Elution Buffer

- **STAR BEADS Tip Comb Plate**, SBTP196,1X96PF, Lot PA27F-XP

**Storage:** 15-30°C

**Expiry Date:** 12/2022

TEST	SPECIFICATIONS	LOT RESULT
Nucleic acid extraction and qRT-PCR:	Average CV $\leq 10.00\%$ (sample Ct vs reference Ct)	6.61%

**Document Release Date:** 09/02/2022

**Document Revision:** 00

**Approved:**



**Quality Control Technician**

**Reviewed:**



**Production Manager**

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.

**CYANAGEN SRL -**

C.F./P.IVA: 02360111203 iscritta nel Registro Imprese di Bologna n. REA: 433637

Capitale sociale sottoscritto e versato € 10.000,00.

"Società soggetta a direzione e coordinamento da parte di Finance & Technology Srl – art.2497 bis cc".