

CERTIFICATE OF ANALYSIS

Product No.

SBK186

Lot No.

NN15D-XP

STAR BEADS Viral DNA/RNA Extraction Kit

Comments: Kit designed for the rapid and efficient isolation of viral DNA and RNA from serum, swabs, plasma, saliva, and other body fluids

Components: STAR BEADS Viral Lysis Buffer, SBLB187, Lot NM09A-LB-2
STAR BEADS Magnetic Beads, SBB188, Lot NM24B-GR-1
STAR BEADS Washing Buffer 1 (concentrate), SBWB189, Lot NF22A-LB-3
STAR BEADS Washing Buffer 2 (concentrate), SBWB190, Lot NM25D-LB-2

Storage: 15-30°C

Expiry Date: 2022/06

TEST	SPECIFICATIONS	LOT RESULT
Nucleic acid extraction and qRT-PCR:	CV≤10.00% (sample Ct vs reference Ct)	2.97

Document Release Date: 04/01/2021


Document Revision: 00

Approved:



Quality Control Technician

Reviewed:



Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.