

CERTIFICATE OF ANALYSIS

Product No.

SBK186,2x32PF

Lot No.

OB01A-XP

STAR BEADS Viral DNA/RNA Extraction Kit Prefilled Plates

Comments: Kit designed for the rapid and efficient isolation of viral DNA and RNA from serum, swabs, plasma, saliva, and other body fluids

Components: STAR BEADS Viral Extraction Plate (x4), SBK197,1x16PF Lot OB01B-XP
STAR BEADS Rod's Tip (1 bag with 8 Rod's Tip), SBK198,2x32 Lot OB01C-XP

Storage: 15-30°C

Expiry Date: 05/2021

TEST	SPECIFICATIONS	LOT RESULT
Nucleic acid extraction and qRT-PCR:	CV ≤10.00% (sample Ct vs reference Ct)	1.12%

Document Release Date: 03/02/2021

Document Revision: 00

Approved:



Quality Control Technician

Reviewed:



Production Manager

Cyanagen warrants that its products conform to the information contained in this and other Cyanagen publications. Purchaser must determine the suitability of the product for its particular use. The values given on the 'Certificate of Analysis' are the results determined at the time of analysis.